


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1884, Rack A, Plate 1, Well B03, 457244  
 [9C24324081D904E0] Aug 13, 2025 11:11:43 Pressure = 60 bar (60 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.413	F 0.49	1.0%	14237	
2	0.500	F 0.60	1.3%	18744	
3	0.712	F 0.85	17.5%	245704	Acetylated F peak
4	0.852	F 1.02	72.8%	1023760	Consistent with F
5	1.212	A 1.01	7.3%	103169	A peak
				Total Area: 1405614	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range