


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2200, Rack A, Plate 1, Well A09, 487135
[992CC26DCE0044CA] Jul 17, 2026 14:32:08 Pressure = 82 bar (82 to 85)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.525	F	0.63	1.2%	7619	
2	0.690	F	0.83	15.6%	97899	Acetylated F peak
3	0.837	F	1.01	74.7%	468430	Consistent with F
4	1.158	A	1.01	8.5%	53150	A peak
Total Area: 627098						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range