


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2140, Rack A, Plate 1, Well E06, 173259
[C8371A6ACE3C8C68] May 06, 2026 13:16:53 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.422	F 0.49	1.0%	11602	
2	0.512	F 0.60	0.4%	5260	
3	0.713	F 0.83	12.5%	148799	Acetylated F peak
4	0.865	F 1.01	62.7%	744407	Consistent with F
5	1.227	A 1.00	23.3%	276527	A peak - REVIEW
Total Area:				1186595	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range