


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1965, Rack A, Plate 1, Well A11, 797010
 [9C22F0870A72ACCA] Oct 31, 2025 13:58:32 Pressure = 68 bar (65 to 68)

AF



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.385	F 0.19	4.1%	79092	
2	0.532	F 0.27	3.0%	57633	
3	0.907	F 0.45	13.0%	252042	
4	1.288	F 0.64	0.6%	12376	
5	1.635	F 0.82	0.8%	14998	
6	1.947	F 0.97	25.6%	497206	Consistent with F
7	2.715	F 1.36	3.5%	68167	
8	2.892	A 0.77	0.2%	3497	
9	2.985	A 0.79	0.6%	11893	
10	3.752	A 1.00	47.6%	925740	A peak - REVIEW
11	4.547	S 0.86	1.1%	21089	3
				Total Area: 1943733	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range