


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well D10, 503345
[D83CC01FBC5D9195] Jul 09, 2026 11:52:11 Pressure = 81 bar (81 to 83)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.517	F	0.62	1.4%	11985	
2	0.692	F	0.84	12.4%	108717	Acetylated F peak
3	0.823	F	0.99	65.7%	574739	Consistent with F
4	1.158	A	1.01	20.6%	179973	A peak - REVIEW
Total Area:					875414	

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range