


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2179, Rack A, Plate 1, Well G04, 486158  
 [993DFA94281A64E0] Jun 18, 2026 15:09:35 Pressure = 70 bar (70 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.432	F 0.51	0.4%	6712	
2	0.515	F 0.61	0.2%	3671	
3	0.710	F 0.85	12.6%	190076	Acetylated F peak
4	0.822	F 0.98	75.5%	1135135	Consistent with F
5	1.197	A 1.01	11.2%	168389	A peak
Total Area:				1503983	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range