


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2137, Rack A, Plate 1, Well B02, 503254
[D8675EEACFC351BF] Apr 30, 2026 11:58:36 Pressure = 63 bar (62 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.427	F 0.50	1.1%	8001	
2	0.518	F 0.61	0.8%	6170	
3	0.712	F 0.83	13.8%	104531	Acetylated F peak
4	0.853	F 1.00	65.4%	494110	Consistent with F
5	1.222	A 1.01	18.8%	142327	A peak
Total Area:				755139	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range