


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2182, Rack A, Plate 1, Well D07, 503330  
[CC6CEA2DC438E448] Jun 23, 2026 11:36:01 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.428	F 0.50	0.7%	7811	
2	0.513	F 0.60	0.4%	4884	
3	0.698	F 0.82	14.0%	167110	Acetylated F peak
4	0.832	F 0.98	72.0%	861217	Consistent with F
5	1.175	A 1.00	13.0%	155489	A peak
Total Area:				1196511	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range