


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1795, Rack A, Plate 1, Well C06, 455878  
 [D934B895D5B2539F] May 14, 2025 11:00:40 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.400	F 0.47	2.2%	23820	
2	0.488	F 0.57	1.2%	13671	
3	0.733	F 0.86	13.7%	151886	Acetylated F peak
4	0.855	F 1.00	70.4%	778667	Consistent with F
5	1.233	A 1.01	12.5%	137662	A peak
Total Area:				1105706	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range