


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2121, Rack A, Plate 1, Well H03, 172987
[997253565429F1B7] Apr 13, 2026 13:29:18 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.47	2.1%	13459	
2	0.505	F 0.58	2.7%	17878	
3	0.718	F 0.82	9.1%	59278	Acetylated F peak
4	0.875	F 1.00	48.9%	319912	Consistent with F
5	1.095	A 0.88	0.7%	4853	
6	1.270	A 1.02	36.5%	238292	A peak - REVIEW
Total Area:				653672	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range