


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2098, Rack A, Plate 1, Well C01, 498828
[9C333F0156BB531F] Mar 19, 2026 12:02:24 Pressure = 60 bar (60 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.430	F 0.50	0.9%	11779	
2	0.520	F 0.60	0.7%	9735	
3	0.720	F 0.83	16.9%	229253	Acetylated F peak
4	0.857	F 0.99	71.2%	968165	Consistent with F
5	1.243	A 1.01	10.4%	141087	A peak

Total Area: 1360019

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range