


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1380, Rack A, Plate 1, Well A03, 27461
[CDC8A74EE3D2C2C8] Apr 17, 2024 10:28:41 Pressure = 79 bar (79 to 79)

FA



| PEAK | RT | REL RT | % CONC | AREA | COMMENT |
|-------------|-------|--------|--------|--------|-------------------|
| 1 | 0.393 | F 0.46 | 1.4% | 8502 | |
| 2 | 0.482 | F 0.57 | 0.6% | 3646 | |
| 3 | 0.727 | F 0.86 | 15.2% | 91353 | Acetylated F peak |
| 4 | 0.867 | F 1.02 | 58.4% | 349954 | Consistent with F |
| 5 | 1.215 | A 1.01 | 24.3% | 145716 | A peak - REVIEW |
| Total Area: | | | | 599171 | |

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range