


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1988, Rack A, Plate 1, Well A09, 171106  
 [8D23ECB5DAC55335] Nov 21, 2025 10:54:15 Pressure = 70 bar (69 to 72)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	0.8%	7835	
2	0.498	F 0.58	0.5%	5204	
3	0.707	F 0.83	13.0%	134597	Acetylated F peak
4	0.847	F 0.99	63.6%	660911	Consistent with F
5	1.015	F 1.19	1.5%	15369	
6	1.238	A 1.02	20.7%	214752	A peak - REVIEW
Total Area:				1038668	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range