


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2156, Rack A, Plate 1, Well G05, 501584
[DD6CB6D2F7F2C440] May 26, 2026 14:04:19 Pressure = 65 bar (65 to 67)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.512	F 0.59	2.8%	20610	
2	0.710	F 0.82	8.1%	59291	Acetylated F peak
3	0.858	F 0.99	50.3%	366421	Consistent with F
4	1.220	A 1.01	38.8%	282505	A peak - REVIEW
Total Area:				728827	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range