


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2038, Rack A, Plate 1, Well E05, 171818  
 [CD3688F650C886E2] Jan 09, 2026 12:23:58 Pressure = 79 bar (78 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.48	1.3%	8713	
2	0.490	F 0.58	0.6%	3643	
3	0.742	F 0.88	10.6%	69195	Acetylated F peak
4	0.843	F 1.00	56.5%	368595	Consistent with F
5	1.012	F 1.20	0.8%	5263	
6	1.220	A 1.02	29.9%	195191	A peak - REVIEW
7	1.383	S 0.89	0.2%	1532	
Total Area:				652132	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range