


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2185, Rack A, Plate 1, Well G03, 486235
[9C6CAB5674B066EA] Jun 25, 2026 14:20:12 Pressure = 72 bar (71 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.49	0.4%	4565	
2	0.517	F 0.61	0.2%	2296	
3	0.688	F 0.81	12.2%	130810	Acetylated F peak
4	0.832	F 0.98	74.5%	801721	Consistent with F
5	1.193	A 1.01	12.7%	136933	A peak
Total Area:				1076325	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range