


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1795, Rack A, Plate 1, Well C04, 457133
 [8C35F895D5BAD1BF] May 14, 2025 10:54:36 Pressure = 58 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.403	F 0.47	1.5%	15410	
2	0.492	F 0.58	1.0%	10474	
3	0.735	F 0.86	18.2%	186002	Acetylated F peak
4	0.858	F 1.01	67.8%	692818	Consistent with F
5	1.228	A 1.01	11.4%	116430	A peak
Total Area:				1021134	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range