


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2061, Rack A, Plate 1, Well C03, 483334  
[CC339B10A96A0CE0] Jan 30, 2026 10:54:07 Pressure = 90 bar (89 to 92)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.502	F 0.60	1.1%	19064	
2	0.698	F 0.83	16.4%	282419	Acetylated F peak
3	0.828	F 0.99	68.1%	1171991	Consistent with F
4	1.193	A 1.01	14.4%	247619	A peak
Total Area:				1721093	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range