


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1359, Rack A, Plate 1, Well G06, 27497
 [9D98F0038BA71F97] Apr 03, 2024 12:12:42 Pressure = 59 bar (59 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.415	F 0.47	1.6%	19773	
2	0.488	F 0.55	3.0%	37362	
3	0.745	F 0.84	11.8%	148688	Acetylated F peak
4	0.873	F 0.98	72.2%	910293	Consistent with F
5	1.252	A 1.00	11.5%	145386	A peak
Total Area:				1261502	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range