


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2008, Rack A, Plate 1, Well B05, 483947
 [C967EB642EB88CC0] Dec 09, 2025 11:26:03 Pressure = 74 bar (73 to 75)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	0.8%	7652	
2	0.500	F 0.59	0.6%	5391	
3	0.708	F 0.83	15.8%	151490	Acetylated F peak
4	0.853	F 1.00	71.4%	682600	Consistent with F
5	1.233	A 1.02	11.4%	109009	A peak
Total Area:				956142	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range