


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1889, Rack A, Plate 1, Well D06, 169742  
 [8C247C894A46DB97] Aug 20, 2025 12:14:50 Pressure = 59 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.405	F 0.48	1.1%	11768	
2	0.493	F 0.58	1.1%	11202	
3	0.700	F 0.83	15.3%	161920	Acetylated F peak
4	0.838	F 0.99	71.8%	762273	Consistent with F
5	1.218	A 1.01	10.7%	113974	A peak
Total Area:				1061137	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range