


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2022, Rack A, Plate 1, Well H03, 482244
[887780C2DD008E40] Dec 19, 2025 13:03:47 Pressure = 73 bar (73 to 75)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.397	F 0.46	1.4%	16345	
2	0.498	F 0.58	0.9%	10162	
3	0.718	F 0.84	15.0%	174945	Acetylated F peak
4	0.867	F 1.01	66.9%	778050	Consistent with F
5	1.223	A 1.01	15.8%	183469	A peak
Total Area:				1162971	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range