


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2083, Rack A, Plate 1, Well H03, 490054
[D86768B2F8E4246A] Feb 27, 2026 13:16:20 Pressure = 52 bar (52 to 54)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.47	0.8%	14398	
2	0.512	F 0.57	0.7%	12514	
3	0.738	F 0.82	13.7%	261567	Acetylated F peak
4	0.887	F 0.99	68.1%	1299980	Consistent with F
5	1.275	A 1.01	16.8%	321699	A peak
Total Area:				1910158	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range