


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2077, Rack A, Plate 1, Well D07, 489344  
[8C6323D2F779531D] Feb 18, 2026 11:13:21 Pressure = 56 bar (55 to 57)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.420	F 0.48	0.9%	21386	
2	0.513	F 0.58	0.9%	22353	
3	0.727	F 0.83	18.0%	450388	Acetylated F peak
4	0.860	F 0.98	67.9%	1701805	Consistent with F
5	1.252	A 0.99	12.4%	309896	A peak
Total Area:				2505828	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range