


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2038, Rack A, Plate 1, Well E03, 488959  
[8D6388F650CA84C0] Jan 09, 2026 12:19:55 Pressure = 79 bar (79 to 81)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.412	F 0.49	1.2%	12297	
2	0.493	F 0.58	0.6%	5836	
3	0.697	F 0.83	12.2%	122133	Acetylated F peak
4	0.837	F 0.99	57.3%	572204	Consistent with F
5	0.990	F 1.17	1.5%	15180	
6	1.210	A 1.01	27.1%	270169	A peak - REVIEW
Total Area:				997819	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range