


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1911, Rack A, Plate 1, Well F05, 170128  
 [D861557E4EDEF93F] Sep 17, 2025 14:15:32 Pressure = 59 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.498	F 0.59	1.3%	21714	
2	0.697	F 0.83	12.4%	202240	Acetylated F peak
3	0.827	F 0.98	61.4%	1005202	Consistent with F
4	0.972	F 1.16	1.2%	19739	
5	1.203	A 1.01	23.3%	380512	A peak - REVIEW
6	1.355	A 1.14	0.4%	6920	
Total Area:				1636327	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range