


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2127, Rack A, Plate 1, Well C12, 498274
[9837189F9DB626CA] Apr 21, 2026 10:57:28 Pressure = 63 bar (62 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.430	F 0.50	0.8%	12086	
2	0.515	F 0.59	0.5%	7611	
3	0.710	F 0.82	16.6%	249280	Acetylated F peak
4	0.847	F 0.98	73.7%	1103525	Consistent with F
5	1.242	A 1.00	8.3%	124717	A peak
Total Area:				1497219	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range