


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1359, Rack A, Plate 1, Well E12, 27428
[DDC9E4038B2D9F15] Apr 03, 2024 11:36:19 Pressure = 60 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.482	F 0.54	5.2%	65322	
2	0.735	F 0.83	9.0%	114386	Acetylated F peak
3	0.885	F 0.99	64.3%	813725	Consistent with F
4	1.252	A 1.00	21.6%	273051	A peak - REVIEW
Total Area:				1266484	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range