


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2124, Rack A, Plate 1, Well E02, 502337
 [98321CB65841D93D] Apr 16, 2026 11:51:33 Pressure = 62 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.432	F 0.49	1.6%	22955	
2	0.505	F 0.57	2.5%	36790	
3	0.722	F 0.82	13.6%	196670	Acetylated F peak
4	0.862	F 0.98	65.9%	951271	Consistent with F
5	1.250	A 1.01	16.4%	236661	A peak
Total Area:				1444347	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range