


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2140, Rack A, Plate 1, Well E08, 173261
[8C675A6ACE368ECA] May 06, 2026 13:20:56 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.428	F 0.50	0.9%	10266	
2	0.517	F 0.60	0.6%	6852	
3	0.705	F 0.82	14.8%	163937	Acetylated F peak
4	0.852	F 0.99	60.8%	675523	Consistent with F
5	1.230	A 1.00	22.9%	254778	A peak - REVIEW

Total Area: 1111356

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range