


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1842, Rack A, Plate 1, Well E02, 458016
 [9D208741548AD39F] Jun 25, 2025 12:58:00 Pressure = 61 bar (61 to 63)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	1.4%	13964	
2	0.495	F 0.58	1.6%	15252	
3	0.710	F 0.83	12.6%	123693	Acetylated F peak
4	0.838	F 0.98	76.4%	749368	Consistent with F
5	1.228	A 1.02	8.0%	78870	A peak
Total Area:				981147	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range