


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 1897, Rack A, Plate 1, Well A04, 169889
[CD302EF17831AEE8] Aug 28, 2025 08:21:17 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.403	F	0.48	0.9%	8242	
2	0.493	F	0.59	0.6%	5885	
3	0.722	F	0.86	12.8%	116708	Acetylated F peak
4	0.830	F	0.99	74.0%	672118	Consistent with F
5	1.215	A	1.01	11.6%	105717	A peak
Total Area: 908670						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range