


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2092, Rack A, Plate 1, Well B12, 492128
 [9C376EF787460C4A] Mar 11, 2026 12:10:40 Pressure = 57 bar (57 to 59)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.432	F 0.49	0.9%	9477	
2	0.515	F 0.59	0.8%	8592	
3	0.723	F 0.83	18.2%	187689	Acetylated F peak
4	0.863	F 0.99	69.9%	722362	Consistent with F
5	1.248	A 1.01	10.2%	105610	A peak
Total Area:				1033730	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range