


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2025, Rack A, Plate 1, Well H02, 490127
[8927C578B32226C8] Dec 23, 2025 13:58:49 Pressure = 75 bar (74 to 77)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.46	0.9%	9796	
2	0.498	F 0.57	0.6%	6476	
3	0.712	F 0.82	14.6%	163565	Acetylated F peak
4	0.848	F 0.98	73.3%	822295	Consistent with F
5	1.230	A 1.01	10.7%	120348	A peak
Total Area:				1122480	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range