


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2185, Rack A, Plate 1, Well G07, 486242
[C82DEF12DE186C48] Jun 25, 2026 14:28:17 Pressure = 72 bar (72 to 74)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.505	F 0.59	1.2%	16228	
2	0.692	F 0.81	9.2%	129708	Acetylated F peak
3	0.833	F 0.98	57.2%	804325	Consistent with F
4	1.182	A 1.00	32.4%	456115	A peak - REVIEW
Total Area:				1406376	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range