


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2054, Rack A, Plate 1, Well F01, 488994  
 [D863CA8AC29F5335] Jan 26, 2026 12:45:11 Pressure = 87 bar (86 to 89)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.422	F 0.49	1.5%	14470	
2	0.500	F 0.58	0.7%	7202	
3	0.707	F 0.82	20.2%	199012	
4	0.830	F 0.97	64.9%	640078	Consistent with F
5	0.988	F 1.15	1.1%	10512	
6	1.200	A 1.01	11.7%	115731	A peak
Total Area:				987005	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range