


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2113, Rack A, Plate 1, Well H05, 498290
 [896253C67D74246A] Apr 07, 2026 14:13:56 Pressure = 59 bar (58 to 60)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.425	F 0.47	0.7%	9942	
2	0.518	F 0.57	0.7%	10027	
3	0.723	F 0.80	14.4%	204733	Acetylated F peak
4	0.863	F 0.95	65.0%	923321	Consistent with F
5	1.255	A 1.00	19.2%	272078	A peak
Total Area:				1420101	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range