


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1870, Rack A, Plate 1, Well A09, 457212
 [C82031B0A506F197] Jul 23, 2025 09:34:34 Pressure = 61 bar (61 to 62)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.407	F 0.49	1.0%	10033	
2	0.495	F 0.59	1.5%	14591	
3	0.703	F 0.84	16.9%	161224	Acetylated F peak
4	0.840	F 1.01	68.3%	653301	Consistent with F
5	1.215	A 1.01	12.2%	116813	A peak
Total Area:				955962	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range