


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2025, Rack A, Plate 1, Well A03, 490106
 [8837D12891A20662] Dec 23, 2025 11:07:02 Pressure = 77 bar (77 to 79)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.408	F 0.48	1.3%	5720	
2	0.500	F 0.59	1.1%	4711	
3	0.708	F 0.83	13.5%	58967	Acetylated F peak
4	0.867	F 1.02	58.8%	256461	Consistent with F
5	1.000	F 1.18	2.4%	10621	
6	1.227	A 1.02	22.9%	99786	A peak - REVIEW
Total Area:				436266	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range