


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2186, Rack A, Plate 1, Well A09, 501497  
 [892CFB46DE9ACCCA] Jun 25, 2026 14:59:20 Pressure = 73 bar (71 to 74)

AF



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.413	F	0.22	2.7%	29382	
2	0.545	F	0.29	2.9%	31739	
3	0.840	F	0.45	5.3%	58420	
4	1.152	F	0.62	0.7%	7136	
5	1.267	F	0.68	0.6%	6316	
6	1.847	F	1.00	21.6%	236414	Consistent with F 3
7	2.608	F	1.41	6.4%	70255	
8	3.138	A	0.88	0.2%	2429	
9	3.565	A	1.00	58.1%	635591	A peak - REVIEW 3
10	4.385	S	0.86	1.5%	16155	
Total Area:					1093837	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range