


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2194, Rack A, Plate 1, Well G08, 493091
[883DD50AE3026EE0] Jul 09, 2026 13:02:55 Pressure = 79 bar (79 to 82)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.523	F	0.63	0.9%	17797	
2	0.688	F	0.83	12.1%	235449	Acetylated F peak
3	0.827	F	1.00	63.6%	1241204	Consistent with F
4	1.143	A	0.99	23.4%	457735	A peak - REVIEW
Total Area: 1952185						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range