


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 1380, Rack A, Plate 1, Well B04, 65079  
[888CB31BB40FBD1D] Apr 17, 2024 10:54:57

Pressure = 83 bar (82 to 83) FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.392	F 0.46	0.9%	13610	
2	0.482	F 0.57	0.6%	9094	
3	0.748	F 0.88	15.3%	233636	Acetylated F peak
4	0.858	F 1.01	59.1%	901737	Consistent with F
5	1.220	A 1.01	24.1%	368445	A peak - REVIEW

Total Area: 1526522

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range