


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2185, Rack A, Plate 1, Well A12, 486210
[9D38BA56761AC648] Jun 25, 2026 12:08:49 Pressure = 72 bar (71 to 73)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.422	F	0.50	1.1%	6578	
2	0.505	F	0.60	0.5%	3027	
3	0.690	F	0.82	10.3%	60233	Acetylated F peak
4	0.845	F	1.01	53.8%	315851	Consistent with F
5	1.187	A	1.01	34.3%	201072	A peak - REVIEW
Total Area: 586761						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range