


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 2156, Rack A, Plate 1, Well B12, 501880
 [CC3CB7C7A82F9337] May 26, 2026 12:13:10 Pressure = 66 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.423	F 0.50	2.7%	11375	
2	0.505	F 0.59	4.2%	17771	
3	0.725	F 0.85	15.6%	66637	
4	0.857	F 1.00	48.8%	208122	Consistent with F
5	1.033	A 0.86	3.6%	15408	
6	1.227	A 1.02	25.2%	107450	A peak - REVIEW
Total Area:				426763	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range