


LABORATORIO PREGEN
 Carrera 15a No 106-42
 BOGOTA

Batch 1974, Rack A, Plate 1, Well A09, 170956
 [8C27B2BB3205DB3F] Nov 11, 2025 11:25:40 Pressure = 67 bar (66 to 68)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.46	1.2%	15662	
2	0.497	F 0.57	0.5%	5962	
3	0.722	F 0.83	13.5%	171406	Acetylated F peak
4	0.852	F 0.98	61.3%	779028	Consistent with F
5	1.018	F 1.18	1.4%	17927	
6	1.242	A 1.01	22.0%	279987	A peak - REVIEW
Total Area:				1269972	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range