


LABORATORIO PREGEN  
Carrera 15a No 106-42  
BOGOTA

Batch 2166, Rack A, Plate 1, Well B03, 501591  
[9D2CB758B4251BB5] Jun 02, 2026 10:30:23 Pressure = 64 bar (63 to 66)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.427	F 0.51	0.4%	5614	
2	0.515	F 0.62	0.3%	4502	
3	0.690	F 0.83	11.0%	148991	Acetylated F peak
4	0.830	F 0.99	65.8%	894393	Consistent with F
5	1.190	A 1.01	22.5%	306200	A peak - REVIEW
Total Area:				1359700	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range