


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 2098, Rack A, Plate 1, Well E12, 498838  
 [C8223F5401EC8662] Mar 19, 2026 13:17:10 Pressure = 59 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.422	F 0.48	0.7%	8725	
2	0.520	F 0.59	0.7%	8692	
3	0.720	F 0.82	16.4%	209809	Acetylated F peak
4	0.857	F 0.98	69.6%	890277	Consistent with F
5	1.242	A 1.00	12.6%	161705	A peak
Total Area:				1279208	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range