


- - - - -
`LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2018, Rack A, Plate 1, Well A08, 171461
[DD2281B40DA02EC0] Dec 15, 2025 09:46:35 Pressure = 76 bar (75 to 77)

FA



PEAK	RT	REL	RT	% CONC	AREA	COMMENT
1	0.395	F	0.46	2.1%	10771	
2	0.492	F	0.58	0.9%	4331	
3	0.708	F	0.83	11.3%	56607	Acetylated F peak
4	0.873	F	1.03	50.3%	253177	Consistent with F
5	1.008	F	1.19	2.1%	10667	
6	1.227	A	1.01	33.3%	167293	A peak - REVIEW
Total Area: 502846						

Codes: 1) Wide A peak
2) Area of A peak < 80%
3) Peak area greater than expected
4) Peak after A2
5) A1c > 10%
6) HbF or variant present
7) Total sample area too small/big
8) A2 is not within normal range