


LABORATORIO PREGEN
Carrera 15a No 106-42
BOGOTA

Batch 2140, Rack A, Plate 1, Well E07, 173260
[C9360F3F9B635BB7] May 06, 2026 13:18:54 Pressure = 63 bar (63 to 65)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.517	F 0.60	1.6%	23203	
2	0.710	F 0.83	11.8%	167033	Acetylated F peak
3	0.848	F 0.99	53.2%	754465	Consistent with F
4	1.228	A 1.00	33.4%	474398	A peak - REVIEW
Total Area:				1419099	

- Codes:
- 1) Wide A peak
 - 2) Area of A peak < 80%
 - 3) Peak area greater than expected
 - 4) Peak after A2
 - 5) Alc > 10%
 - 6) HbF or variant present
 - 7) Total sample area too small/big
 - 8) A2 is not within normal range