


LABORATORIO PREGEN  
 Carrera 15a No 106-42  
 BOGOTA

Batch 1827, Rack A, Plate 1, Well A09, 456684  
 [DC358422D902F3BD] Jun 11, 2025 09:30:05 Pressure = 60 bar (59 to 61)

FA



PEAK	RT	REL RT	% CONC	AREA	COMMENT
1	0.402	F 0.47	1.3%	15375	
2	0.488	F 0.57	1.3%	15644	
3	0.715	F 0.84	11.8%	142537	Acetylated F peak
4	0.842	F 0.99	69.4%	839212	Consistent with F
5	1.233	A 1.01	16.2%	196184	A peak
Total Area:				1208952	

- Codes:
- 1) Wide A peak
  - 2) Area of A peak < 80%
  - 3) Peak area greater than expected
  - 4) Peak after A2
  - 5) Alc > 10%
  - 6) HbF or variant present
  - 7) Total sample area too small/big
  - 8) A2 is not within normal range